

Gas detector safety alert

Safety Flash Published on 27 October 2011 Generated on 17 January 2025 IMCA SF 11/11

The UK Health and Safety Executive (HSE) has issued an alert relating to gas detectors, following its investigations into the Status Scientific Controls portable gas detector type Mentor PGD2.

It has been discovered that in certain circumstances these instruments could continue to remain active when there is insufficient battery power for them to work accurately.

Further information can be found at hse.gov.uk/safetybulletins/gasdetectors.htm

IMCA Safety Flashes summarise key safety matters and incidents, allowing lessons to be more easily learnt for the benefit of the entire offshore industry.

The effectiveness of the IMCA Safety Flash system depends on the industry sharing information and so avoiding repeat incidents. Incidents are classified according to IOGP's Life Saving Rules.

All information is anonymised or sanitised, as appropriate, and warnings for graphic content included where possible.

IMCA makes every effort to ensure both the accuracy and reliability of the information shared, but is not be liable for any guidance and/or recommendation and/or statement herein contained.

The information contained in this document does not fulfil or replace any individual's or Member's legal, regulatory or other duties or obligations in respect of their operations. Individuals and Members remain solely responsible for the safe, lawful and proper conduct of their operations.

Share your safety incidents with [IMCA online](#). Sign-up to receive Safety Flashes [straight to your email](#).